

Addendum No 1

Date: 02/12/2019

This is with reference to the tender notification for “**Electrical Signal Generator**”,
Tender Reference No: NITK/CSD/TPEM/2019/ESG/04 Dated: 22/10/2019

Please note the following additions to the technical specifications:

- 1. Signal maximum voltage >70V**
- 2. Maximum continuous current rating >20A**
- 3. Frequency response - 1MHz**
- 4. Slew rate – 100V/micro sec**

**Sd/-
Buyer
(Dr.Navin Karanth P)**

**Sd/-
Coordinator
Centre for System Design**