


NATIONAL INSTITUTE OF TECHNOLOGY KARNATAKA, SURATHKAL

DEPARTMENT OF HEALTH CARE CENTRE
POST SRINIVASNAGAR, MANGALORE – 575 025 (D K)

Phone: (0824) 2474000.
E- mail: info@nitk.ac.in

Fax: (0824) 2474033
Website: http://www.nitk.ac.in



NOTICE INVITING QUOTATION

Notification. No: NITK/HCC/Analyzer/2021-22/A8.....

dated: 15/07/2021

Name of Goods	BIOCHEMISTRY ANALYZER (SEMI AUTOMATIC)
Estimated Amount:	1,75,000/-
Time for Supply of item after release of Purchase order	15 days
Document Download / Sale Start Date	15/07/2021 at 5:30 PM
Clarification Start Date	15/07/2021 at 5:30 PM
Clarification End Date	23/07/2021 at 3:30 PM
Bid Submission Start Date	15/07/2021 at 5:30 PM
Last Date for submission of bids	30/07/2021 before 3.00 PM
Bid Opening Date	30/07/2021 before 4.00 PM (if possible)
Address for Submission of bids [Type Address of the Department and Contact Person Details with Phone Number]	Superintendent ,(Purchase) Purchase Section NITK, Surathkal P O Srinivasnagar – 575 025 Ph :0824 -2473014



NATIONAL INSTITUTE OF TECHNOLOGY KARNATAKA, SURATHKAL
DEPARTMENT OF HEALTH CARE CENTRE
POST SRINIVASNAGAR, MANGALORE – 575 025

Phone: (0824) 2474000
E- mail: info@nitk.ac.in

Fax: (0824) 2474033
Website: <http://www.nitk.ac.in>

Notification. No: NITK/HCC/Analyzer/2021-22/A8

dated: 15/07/2021

NOTICE INVITING QUOTATION (NIQ)

The National Institute of Technology Karnataka, Surathkal (in short – NITK, Surathkal) is an Institute Of National Importance Under Ministry of Education Govt of India, imparting Technical Education and engaged in Research Activities. It is proposed to procure the items for the departmental academic/research activities.

Sealed Quotations as per the Price Schedule given in this NIQ are invited for the following items subject to the terms and conditions, from the reputed manufacturers or its authorised dealers so as to reach on or before scheduled date and time. The quotations in the firm's Business letter head should be address to the "Director, NITK, Surathkal". The envelope shall be superscribed with the Quotation Notification Number and the Name of the Goods for which quotation is submitted.

Name of Goods	BIOCHEMISTRY ANALYZER (SEMI AUTOMATIC)
Estimated Amount:	1,75,000/-
Time for Supply of item after release of Purchase order	15 days
Document Download / Sale Start Date	15/07/2021 at 5:30 PM
Clarification Start Date	15/07/2021 at 5:30 PM
Clarification End Date	23/07/2021 at 3:30 PM
Bid Submission Start Date	15/07/2021 at 5:30 PM
Last Date for submission of bids	30/07/2021 before 3.00 PM
Bid Opening Date	30/07/2021 before 4.00 PM (if possible)

Sd/-
Asst Registrar A/cs (Purchase)

Note: Institute shall not be responsible for any postal delay about non-receipt /non-delivery of the bids or due to wrong addressee.

SECTION-1
Terms and Conditions

1. The rates should be quoted for preferably FOR destination from supply within India.
2. The bidder shall indicate the excise duty exemption for the goods if applicable.
3. The rate quoted should be on unit basis. Taxes and other charges should be quoted separately, considering exemptions if any. The rate should be quoted in INR only
4. Rate quoted should be inclusive of Testing, commissioning and Installation of equipment and Training.
5. Payment: No advance payment will be made. Payment will be made only after the supply of the item in good and satisfactory condition and receipt of performance security by supplier.
6. Guarantee/Warranty period should be specified for the complete period should be specified in section 3 of this tender document.
7. Period requirement for the supply and installation of item should be specified in section 3 of this tender document.
8. In case of dispute, the matter will be subject to Mangalore Jurisdiction only.

SECTION-2
SCHEDULE OF REQUIREMENTS, SPECIFICATIONS AND ALLIED DETAILS
[To be filled up by the Department / Centre of NITK, Surathkal]

Item(s) Name to be Procured : **BIOCHEMISTRY ANALYZER
(SEMI AUTOMATIC)**

Brief Specifications of the Item(s)
(Attach Additional Sheet if necessary) : **Attached**

Quantity : **01 No.**

Any other details / requirement :

Warranty Period required : **2 years**

Delivery Schedule expected
after placement of Purchase order
(in Weeks) : **2 weeks**

SECTION 3
PRICE SCHEDULE

[To be used by the bidder for submission of the quotation]

-
1. Item Name :
 2. Specifications
(Conforming to Schedule of requirements
Enclose additional sheets if necessary) :
 3. Currency and Unit Price :
 4. Quantity :
 5. Item Cost (Sl No. 3 * Sl. No. 4) :
 6. Taxes and Other Charges :
(i) Specify the type of taxes and duties
in percentages and also in figures.
(ii) Specify Other Charges in figures.
 7. **Warranty Period** :
(Conforming to the Schedule of requirements)
 8. Delivery Schedule :
(Conforming to the Schedule of requirements)
 9. Name and address of the Firm for
placing purchase order :
 10. Name and address of Indian authorized
agent (in case of imports only) :

Signature of the Bidder: _____

Name and Designation: _____

Business Address : _____

Place:
Date

Seal of the Bidder's Firm

SECTION 4
CONTRACT FORM

[To be provided by the bidder in the business letter head]

1. (Name of the Supplier's Firm) hereby abide by the delivery schedule mentioned in this document for supply of the items if the purchase order is awarded.
2. The item will be supplied conforming to the specifications stated in this document without any defect and deviations.
3. Warranty will be given for the period mentioned in this document and service will be rendered to the satisfaction of NITK, Surathkal during this period.

Signature of the Bidder: _____

Name : _____

Business Address : _____

Place :

Date :

Seal of the Bidder's Firm

Technical Specifications

Test Programme	: 150 totally OPEN test programs, directly selectable through keypad
Optical System	: Static photometer
Light source	: Halogen lamp 12V/20 Watts
Detector	: 8 silicon UV/VIS photodiode array
Cuvette	: Unique Triple Cuvette System, 18 μ l Flow Cell Cuvette, Rectangular polystyrene Cuvette, round Glass tube for Coagulation
Aspiration Volume	: From 300 μ l to 999 μ l
Zero set	: Automatic
Temp. Control	: Microprocessor controlled peltier element, programmable at 25 °C, 30°C & 37°C & 'OFF'
Memory Storage	: 1000 Results, 2 level monthly Q.C. data, Reagent Blank
Calibration	: Automatic against standard
On-board Self-Test	: Software driven self-diagnosis for easy maintenance
Keyboard	: 24 keys, rugged, water proof, tactile legend membrane panel
Display	: High Resolution Graphic LCD 240 x 64 dots with Backlit view area 134 mm x 40 mm
Printer	: High Resolution 320 dots per line, thermal type with graphics facility
Interface	: USB connectivity to PC, Direct Printer connectivity through USB port & Port for External Keyboard
Warranty	: 2 years