

# NATIONAL INSTITUTE OF TECHNOLOGY KARNATAKA, SURATHKAL


DEPARTMENT OF MECHANICAL ENGINEERING  
POST SRINIVASNAGAR, MANGALORE – 575 025 (D K)

Phone: (0824) 2474000.

Fax: (0824) 2474033

E- mail: [info@nitk.ac.in](mailto:info@nitk.ac.in)

Website: <http://www.nitk.ac.in>



## NOTICE INVITING QUOTATION

Notification. No: NITK/DPD/OH35/ME/4853/2022-23

Dated: 06.03.2023

<b>Name of Goods</b>	Open Cup Flash Point Tester
<b>Estimated Amount:</b>	1,00,000/-
<b>Time for Supply of item after release of Purchase order</b>	<b>Two Weeks</b>
<b>Document Download / Sale Start Date</b>	07.03.2023
<b>Clarification Start Date</b>	07.03.2023
<b>Clarification End Date</b>	14.03.2023
<b>Bid Submission Start Date</b>	07.03.2023
<b>Last Date for submission of bids</b>	21.03.2023 <b>before 3.00 PM</b>
<b>Bid Opening Date</b>	21.03.2023 <b>before 3.00 PM</b>
<b>Address for Submission of bids</b>	Dr. Vasudeva M. Assistant Professor, Department of Mechanical Engineering NITK Surathkal Mangaluru – 575025 9008889796



**NATIONAL INSTITUTE OF TECHNOLOGY KARNATAKA, SURATHKAL**  
**DEPARTMENT OF MECHANICAL ENGINEERING**  
POST SRINIVASNAGAR, MANGALORE – 575 025

Phone: (0824) 2474000  
E- mail: info@nitk.ac.in

Fax: (0824) 2474033  
Website: http://www.nitk.ac.in

Notification No: NITK/DPD/OH35/ME/4853/2022-23

Date: 06.03.2023

**NOTICE INVITING QUOTATION (NIQ)**

The National Institute of Technology Karnataka, Surathkal (in short – NITK, Surathkal) is an Institute Of National Importance Under Ministry of Education Govt of India, imparting Technical Education and engaged in Research Activities. It is proposed to procure the items for the departmental academic/research activities.

Sealed Quotations as per the Price Schedule given in this NIQ are invited for the following items subject to the terms and conditions, from the reputed manufacturers or its authorised dealers so as to reach on or before scheduled date and time. The quotations in the firm's Business letter head should be address to the "Director, NITK, Surathkal". The envelope shall be superscribed with the Quotation Notification Number and the Name of the Goods for which quotation is submitted.

<b>Name of Goods</b>	Open Cup Flash Point Tester
<b>Estimated Amount:</b>	1,00,000/-
<b>Time for Supply of item after release of Purchase order</b>	<b>Two Weeks</b>
<b>Document Download / Sale Start Date</b>	07.03.2023
<b>Clarification Start Date</b>	07.03.2023
<b>Clarification End Date</b>	14.03.2023
<b>Bid Submission Start Date</b>	07.03.2023
<b>Last Date for submission of bids</b>	21.03.2023 <b>before 3.00 PM</b>
<b>Bid Opening Date</b>	21.03.2023 <b>before 3.00 PM</b>

Sd/-  
HOD

Note: Institute shall not be responsible for any postal delay about non-receipt /non-delivery of the bids or due to wrong addressee.

**SECTION-1**  
**Terms and Conditions**

1. The rates should be quoted for preferably FOR destination from supply within India.
2. The bidder shall indicate the excise duty exemption for the goods if applicable.
3. The rate quoted should be on unit basis. Taxes and other charges should be quoted separately, considering exemptions if any. The rate should be quoted in INR only
4. Rate quoted should be inclusive of Testing, commissioning and Installation of equipment and Training.
5. Payment: No advance payment will be made. Payment will be made only after the supply of the item in good and satisfactory condition and receipt of performance security by supplier.
6. Guarantee/Warranty period should be specified for the complete period should be specified in section 3 of this tender document.
7. Period requirement for the supply and installation of item should be specified in section 3 of this tender document.
8. In case of dispute, the matter will be subject to Mangalore Jurisdiction only.

**SECTION-2**  
**SCHEDULE OF REQUIREMENTS, SPECIFICATIONS AND ALLIED DETAILS**  
[ To be filled up by the Department / Centre of NITK, Surathkal ]

---

Item(s) Name to be Procured : Open Cup Flash Point Tester

Brief Specifications of the Item(s)  
(Attach Additional Sheet if necessary) : Additional sheet Attached

Quantity : 01

Any other details / requirement : No

Warranty Period required : 01 Year

Delivery Schedule expected  
after placement of Purchase order  
(in Weeks) : 2 weeks

**SECTION 3  
PRICE SCHEDULE**

[ To be used by the bidder for submission of the quotation]

- 
1. Item Name :
  2. Specifications  
(Conforming to Schedule of requirements  
Enclose additional sheets if necessary) :
  3. Currency and Unit Price :
  4. Quantity :
  5. Item Cost (Sl No. 3 \* Sl. No. 4 ) :
  6. Taxes and Other Charges :  
(i) Specify the type of taxes and duties  
in percentages and also in figures.  
(ii) Specify Other Charges in figures.
  7. **Warranty Period** :  
**(Conforming to the Schedule of requirements)**
  8. Delivery Schedule :  
(Conforming to the Schedule of requirements)
  9. Name and address of the Firm for  
placing purchase order :
  10. Name and address of Indian authorized  
agent ( in case of imports only) :

**Signature of the Bidder:** \_\_\_\_\_

**Name and Designation:** \_\_\_\_\_

**Business Address** : \_\_\_\_\_

\_\_\_\_\_

**Place:**

**Date:**

**Seal of the Bidder's Firm**

**SECTION 4**  
**CONTRACT FORM**

[ To be provided by the bidder in the business letter head]

---

1. (Name of the Supplier's Firm) hereby abide by the delivery schedule mentioned in this document for supply of the items if the purchase order is awarded.
2. The item will be supplied conforming to the specifications stated in this document without any defect and deviations.
3. Warranty will be given for the period mentioned in this document and service will be rendered to the satisfaction of NITK, Surathkal during this period.

**Signature of the Bidder:** \_\_\_\_\_

**Name** : \_\_\_\_\_

**Business Address** : \_\_\_\_\_

\_\_\_\_\_

**Place :**  
**Date :**

**Seal of the Bidder's Firm**

## Specifications for Open Cup Flash Point tester

01.	Type	Cleveland Open cup (semi automatic)
02.	Temperature Indicator	Digital
03.	COC Cup	Brass & chrome plated
04.	Accuracy	0.1 °C
05.	Flame type	Jet type for flame
06.	Ignition type	Gas
07.	Temperature Sensor	PT-100 in stainless steel sheath.
08.	Flash/Fire detection	Manual/Visual detection
09.	Heating energy regulator	Digital Thyristor Regulator
10.	Power supply	230 V, 1 Phase, 50 Hz/60 Hz.
11.	Flash & Fire point range	79-400 °C
12.	Display	Digital display for thyristor regulator Digital display of sample temperature.
13.	Flame-Sampling Arrangement	DC gear motor, vertical movement reciprocating to & fro motion.
14.	Accessories	COC cup, Side ignitor spare coil, LPG tube
15.	Cooling	Forced Cooling at the end of the test.
16.	Warranty	One year onsite warranty